



Sağlık Ürünleri İç ve

One Step HCV Test

(Serum/Plasma)

FOR PROFESSIONAL IN VITRO DIAGNOSTIC USE ONLY

Intended Use

The One Step HCV Test is a rapid chromatographic immunoassay for the qualitative detection of antibodies to Hepatitis C Virus (HCV) in Serum and Plasma to aid in the diagnosis of Hepatitis C Virus infection.

Summary

Hepatitis C Virus now is recognized as a major of chronic hepatitis, transfusion acquired non-A, non-B hepatitis and liver disease throughout the world. HCV is an enveloped positive-sense, single-stranded RNA virus. Clinical diagnostic issues related to HCV is the detection of HCV antibodies in Whole Blood /Serum / Plasma. One step HCV Test is a simple, visual qualitative test that detects antibodies in human Serum and Plasma. The test is based on immunochromatography and can give a result within 15 minutes.

Principle

The One Step HCV Test is a qualitative membrane strip based immunoassay for the detection of HCV antibodies in Serum and Plasma. In this test procedure, recombinant HCV antigen is immobilized in the test line region of the device. After a specimen is placed in the specimen well, it reacts with HCV antigen coated particles that have been applied to the specimen pad. This mixture migrates chromatographically along the length of the test strip and interacts with the immobilized HCV antigen. If the specimen contains HCV antibodies, a colored line will appear in the test line region indicating a positive result. If the specimen does not contain HCV antibodies, a colored line will not appear in this region indicating a negative result. To serve as a procedural control, a colored line will always appear at the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

Storage and Stability

- Store as packaged in the sealed pouch at room temperature or refrigerated (4-30°C or 40-

86°F). The test device is stable through the expiration date printed on the sealed pouch.

- The test must remain in the sealed pouch until use.

Additional Special Equipment

Materials Provided

- Test devices
- Disposable specimen droppers
- Package insert

Materials Required But Not Provided

- Timer
- Centrifuge
- Specimen collection containers

Precautions

- For professional in vitro diagnostic use only. Do not use after expiration date.
- Do not eat, drink or smoke in the area where the specimens and kits are handled.
- Handle all specimens as if they contain infectious agents.
- Observe established precautions against microbiological hazards throughout all procedures and follow the standard procedures for proper disposal of specimens.
- Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Follow standard biosafety guidelines for handling and disposal of potential infective material.
- Humidity and temperature can adversely affect results.

Specimen Collection and Preparation

1. The One Step HCV Test can be performed used on Serum and Plasma.
2. To collect whole blood, serum or plasma specimens following regular clinical laboratory procedures.
3. Separate serum or plasma from blood as soon as possible to avoid hemolysis. Use only clear non-hemolyzed specimens.
4. Testing should be performed immediately after specimen collection. Do not leave the specimens at room temperature for prolonged periods. Serum and plasma specimens may be stored at 2-8°C for up to 3 days. For long term storage, specimens should be kept below



Sağlık Ürünleri İç ve


-20°C.

5. Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly.

Test Procedure

Allow the test, specimen, buffer and/or controls to reach room temperature 15-30°C (59-86°F) prior to testing.

1. Bring the pouch to room temperature before opening it. Remove the test device from the sealed pouch and use it as soon as possible.
2. Place the test device on a clean and level surface.
3. Hold the dropper vertically and transfer 3 drops of Serum or Plasma (approximately 100 µl) to the specimen well(S) of the test device, then start the timer. See illustration below.
4. Wait for the colored line(s) to appear. Read results at 15 minutes. Do not interpret the result after 20 minutes.



Notes:

Applying sufficient amount of specimen is essential for a valid test result. If migration (the wetting of membrane) is not observed in the test window after one minute, add one more drop of specimen (serum or plasma) to the specimen well.

Interpretation of Results

Positive: Two lines appear. One line should always appear in the control line region(C), and

another one apparent colored line should appear in the test line region.

Negative: One colored line appears in the control region(C).No apparent colored line appear in the test line region.

Invalid: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test device. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

Limitations

1. The One Step HCV Test is for in vitro diagnostic use only. The test should be used for the detection of HCV antibodies in Serum or Plasma specimens only. Neither the quantitative value nor the rate of increase in HCV antibodies can be determined by this qualitative test.
2. The One Step HCV Test will only indicate the presence of HCV antibodies in the specimen and should not be used as the sole criteria for the diagnosis of Hepatitis C Virus infection.
3. As with all diagnostic tests, all results must be interpreted together with other clinical information available to the physician.
4. If the test result is negative and clinical symptoms persist, additional testing using other clinical methods is recommended. A negative result does not at any time preclude the possibility of HCV infection.

Manufacturer

Company Name: SPESERA SAĞLIK ÜRÜNLERİ İÇ VE DIŞ TİC. LTD. ŞTİ.

Address: Oruç Reis mah. Giyimkent Sit. 17. Sok. No:23 A Esenler/İSTANBUL

Tel: (0212)438-16-39

Fax: (0212)438-16-40

Email: spesera@hotmail.com